



POCT Nova StatStrip Meter QA Log

Final Approval: Dr Bassam Nassar

Hospital	
Unit/Floor	
Month/Year	
Meter Tag/ Serial #	

Daily Checks	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Controls dated																															
Strips dated																															
Run Controls(every 24hrs)																															
Clean exterior of meter(after each use)																															
Place Meter in Docking Station																															
Check battery																															

Semi-Annual Maintenance

Semi-Annual	March	September
Collect Light Green PST sample to be sent to the lab for comparison with meter.		
Process patient for glucose monitoring testing within <i>one hour</i> of blood work being drawn.		
Record both results on the POC Accuracy Check Form.		



POCT Nova StatStrip Meter QA Log

Final Approval: Dr Bassam Nassar

Retain Daily/Monthly checklist in a binder available for review.

Unscheduled Maintenance

Problems/Sent to Biomed

Date & Time	Initials	Problem and Action Taken	Meter Tag/Serial Number	Date & Time Sent to Biomed	Date & Time Returned from Biomed

Retain Daily/Monthly checklist in a binder available for review.